



Declaration of Conformity

We B. Braun Melsungen AG, Carl-Braun-Straße 1, 34212 Melsungen, Germany declare, that the following products:

- 8714827 Perfusor® compact
- 8700150 Universal clamp MICU

are in compliance with the paragraphs of the following standard:

EN 1789

Medical vehicles and their equipment – Road ambulances

6.4.1 Vibration & Shock Resistance Tests

- Shock resistance -
- Vibration (sinusoidal) -
- Vibration (broad band noise) -

B. Braun Melsungen AG

Melsungen, 2009-10-28

i. V.

Rolf Heitmeier
Director New Development and
Hardware Technology Infusion Pumps

i. V.

Ullrich Trottier
Director Regulatory Affairs CoE AIS